



HelicoMax^{plus}® Test Procedure

This is an in vitro diagnostic medical device for professional use only.


1 Open test




2 Place biopsy on test




3 Re-seal test



4 Squeeze with thumb





5 Wait one hour



6 Test Result

Red = positive

Yellow = negative



WARNING: Potentially Biohazardous Material

Treat all biopsy specimens as if capable of transmitting disease.